

EC - Certification of Conformity

Registration No. AGC09920200402E1

Certificate Holder Hunan Tuogao Medical Technology Co. , Ltd.
Room 310-312, building 4, Science and Technology Industrial Park,
Hunan University, 001 Jinzhou North Road, Ningxiang high-tech industrial park

Product Designation Infrared Thermometer

Brand Name N/A

Model / Series Models TG8818N

Manufacturer Hunan Tuogao Medical Technology Co. , Ltd.
Room 310-312, building 4, Science and Technology Industrial Park,
Hunan University, 001 Jinzhou North Road, Ningxiang high-tech industrial park

Requirement	Applied Standards	Document Evidence	Result
Medical Devices Directive	EN 60601-1:2006+A1:2013+A12:2014	Test Report: AGC09920200402ES01	Conform
	EN 60601-1-2:2015	Test Report: AGC09920200402EE01	Conform



Signed by General Manager (King Zhang)

Issue Date: Apr. 27, 2020



Recognized by Attestation of Global Compliance (Shenzhen) Co., Ltd., in accordance with the Medical Devices Directive 93/42/EEC.

The certificate doesn't imply assessment of the production. The Applicant of the certificate is authorized to use this certificate in connection with EC declaration of conformity to the Directive. The certificate is only applicable to the equipments described above. This certificate shall not be re-produced except in full without the written approval of Attestation of Global Compliance (Shenzhen) Co., Ltd.

Note: This certificate is part of the full test report(s) and should be used in conjunction with it.

Version: 2.1